



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005**

**PL Medical Co., LLC**  
321 Ellis Street  
New Britain, CT 06051-3504  
Mr. Sushil K. Kanwar  
Phone: 860-223-8882 Fax: 860-223-5941  
E-Mail: [info@plmedical.com](mailto:info@plmedical.com)  
URL: <http://www.plmedical.com>

**IONIZING RADIATION DOSIMETRY**

**NVLAP LAB CODE 200666-0**

***Scope of Accreditation:***

This facility has been evaluated and deemed competent to process the radiation dosimeter listed below through employing a Panasonic Automatic reader model UD 710A.

This facility is accredited to process the following dosimeter by virtue of actual demonstration of compliance with ANSI/HPS N13.11-2001 through testing.

Panasonic TLD model UD-802 in a 802-1 PL Medical/Ribbel holder for ANSI/HPS N13.11-2001 categories IA, IIA, IIIA, IVA, VAA.

2009-10-01 through 2010-09-30

*Effective dates*

*For the National Institute of Standards and Technology*